

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/01/2022

OPERATOR NAME:  
ROBERT M FOO  
Permit #: 46059  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:40
INTERNAL STANDARD	VERIFIED	03:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:41
BLANK TEST	0.00	03:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:43
BLANK TEST	0.00	03:44

