

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/24/2022

OPERATOR NAME:  
ROBERT M FOO  
Permit #: 46059  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:05
INTERNAL STANDARD	VERIFIED	00:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09

