

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/31/2022

OPERATOR NAME:  
SAMANTHA M MCCONNELL  
Permit #: 47102  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:29
INTERNAL STANDARD	VERIFIED	00:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33

