

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 11/03/2022

OPERATOR NAME:
 TYLER R MELLO
 Permit #: 46781
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20

