

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/07/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:10
INTERNAL STANDARD	VERIFIED	02:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:10
BLANK TEST	0.00	02:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:14

