

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/14/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:47
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:48
BLANK TEST	0.00	02:49

