

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/21/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:04

