

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/05/2022

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:06
INTERNAL STANDARD	VERIFIED	05:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:06
BLANK TEST	0.00	05:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:07
BLANK TEST	0.00	05:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:08
BLANK TEST	0.00	05:09

