

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/19/2022

OPERATOR NAME:
JEFFREY S DAIGLE
Permit #: 41298
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:25
INTERNAL STANDARD	VERIFIED	04:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:26
BLANK TEST	0.00	04:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:27
BLANK TEST	0.00	04:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:28
BLANK TEST	0.00	04:29

