

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/27/2022

OPERATOR NAME:
KIRSTEN M LUNDER
Permit #: 44441
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:25
INTERNAL STANDARD	VERIFIED	06:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:27
BLANK TEST	0.00	06:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:28
BLANK TEST	0.00	06:29

