

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/09/2023

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:46
INTERNAL STANDARD	VERIFIED	01:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:49
BLANK TEST	0.00	01:50

