

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/16/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:40
INTERNAL STANDARD	VERIFIED	02:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:43

