

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 02/01/2023

OPERATOR NAME:
 ROBERT M FOO
 Permit #: 46059
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:57
INTERNAL STANDARD	VERIFIED	03:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:57
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:59
BLANK TEST	0.00	04:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:00
BLANK TEST	0.00	04:01

