

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/06/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:36
INTERNAL STANDARD	VERIFIED	01:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39

