

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/27/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:06
INTERNAL STANDARD	VERIFIED	04:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:06
BLANK TEST	0.00	04:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09

