

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/06/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:28
BLANK TEST	0.00	02:29

