

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/13/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:13

