

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/17/2023

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:49
INTERNAL STANDARD	VERIFIED	04:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:52
BLANK TEST	0.00	04:53

