

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 05/08/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:39
INTERNAL STANDARD	VERIFIED	04:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:42
BLANK TEST	0.00	04:43

