

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 05/22/2023

OPERATOR NAME:
 JACOB T SAUNDERS
 Permit #: 51176
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:15
INTERNAL STANDARD	VERIFIED	05:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:16
BLANK TEST	0.00	05:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:17
BLANK TEST	0.00	05:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:18
BLANK TEST	0.00	05:19

