

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 05/29/2023

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:02
INTERNAL STANDARD	VERIFIED	15:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:05
BLANK TEST	0.00	15:06

