

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/02/2023

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:47
INTERNAL STANDARD	VERIFIED	05:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:50
BLANK TEST	0.00	05:51

