

# SUPERVISOR TEST

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 06/05/2023

OPERATOR NAME:  
 JACOB T SAUNDERS  
 Permit #: 51176  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 23090

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 01:33 |
| INTERNAL STANDARD       | VERIFIED | 01:33 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 01:33 |
| BLANK TEST              | 0.00     | 01:34 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 01:34 |
| BLANK TEST              | 0.00     | 01:35 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 01:35 |
| BLANK TEST              | 0.00     | 01:36 |

