

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/12/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:57
INTERNAL STANDARD	VERIFIED	04:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:58
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:59
BLANK TEST	0.00	05:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:00
BLANK TEST	0.00	05:01

