

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 06/20/2023

OPERATOR NAME:
 HUNTER L BARTON
 Permit #: 49353
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:24
INTERNAL STANDARD	VERIFIED	03:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:25
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:26
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:27
BLANK TEST	0.00	03:28

