

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/26/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35

