

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 07/01/2023

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:02
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:05

