

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 07/03/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:51
INTERNAL STANDARD	VERIFIED	01:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:54
BLANK TEST	0.00	01:55

