

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/10/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:39
INTERNAL STANDARD	VERIFIED	05:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:40
BLANK TEST	0.00	05:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43

