

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/17/2023

OPERATOR NAME:
 RYAN S PEDONE
 Permit #: 35878
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:53
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:53
BLANK TEST	0.00	05:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56

