

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/31/2023

OPERATOR NAME:
 HUNTER L BARTON
 Permit #: 49353
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:15
INTERNAL STANDARD	VERIFIED	04:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:16
BLANK TEST	0.00	04:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:17
BLANK TEST	0.00	04:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:18
BLANK TEST	0.00	04:19

