

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/01/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:27
INTERNAL STANDARD	VERIFIED	04:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:27
BLANK TEST	0.00	04:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:28
BLANK TEST	0.00	04:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:29
BLANK TEST	0.00	04:30

