

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/28/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:23
INTERNAL STANDARD	VERIFIED	04:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:23
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:26
BLANK TEST	0.00	04:27

