

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/11/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55

