

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/18/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:30
INTERNAL STANDARD	VERIFIED	04:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:30
BLANK TEST	0.00	04:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:32
BLANK TEST	0.00	04:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:33
BLANK TEST	0.00	04:34

