

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 09/25/2023

OPERATOR NAME:
 JACOB T SAUNDERS
 Permit #: 51176
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:57
INTERNAL STANDARD	VERIFIED	01:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:57
BLANK TEST	0.00	01:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:00
BLANK TEST	0.00	02:01

