

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/01/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:58
INTERNAL STANDARD	VERIFIED	04:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:59
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:00
BLANK TEST	0.00	05:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:01
BLANK TEST	0.00	05:02

