

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/09/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:29
INTERNAL STANDARD	VERIFIED	06:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:30
BLANK TEST	0.00	06:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:31
BLANK TEST	0.00	06:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:32
BLANK TEST	0.00	06:33

