

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 10/16/2023

OPERATOR NAME:
 JACOB T SAUNDERS
 Permit #: 51176
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:22
INTERNAL STANDARD	VERIFIED	04:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:23
BLANK TEST	0.00	04:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:24
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26

