

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 10/30/2023

OPERATOR NAME:
 JACOB T SAUNDERS
 Permit #: 51176
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:03
INTERNAL STANDARD	VERIFIED	05:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:04
BLANK TEST	0.00	05:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:05
BLANK TEST	0.00	05:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:06
BLANK TEST	0.00	05:07

