

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/01/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:49
INTERNAL STANDARD	VERIFIED	04:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:51
BLANK TEST	0.00	04:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:53
BLANK TEST	0.00	04:53

