

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/13/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10

