

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/20/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:03
INTERNAL STANDARD	VERIFIED	05:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:03
BLANK TEST	0.00	05:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:05
BLANK TEST	0.00	05:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:06
BLANK TEST	0.00	05:07

