

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 12/01/2023

OPERATOR NAME:  
NICHOLAS J LEHOISKY  
Permit #: 49359  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:26
INTERNAL STANDARD	VERIFIED	00:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:28
BLANK TEST	0.00	00:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30

