

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 12/04/2023

OPERATOR NAME:  
JACOB T SAUNDERS  
Permit #: 51176  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:27
INTERNAL STANDARD	VERIFIED	05:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:27
BLANK TEST	0.00	05:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:28
BLANK TEST	0.00	05:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:30
BLANK TEST	0.00	05:31

