

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/11/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:31
INTERNAL STANDARD	VERIFIED	04:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:32
BLANK TEST	0.00	04:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:33
BLANK TEST	0.00	04:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:34
BLANK TEST	0.00	04:35

