

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/18/2023

OPERATOR NAME:
JACOB T SAUNDERS
Permit #: 51176
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:18
INTERNAL STANDARD	VERIFIED	04:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:19
BLANK TEST	0.00	04:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:20
BLANK TEST	0.00	04:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:21
BLANK TEST	0.00	04:22

