

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/29/2024

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:47
INTERNAL STANDARD	VERIFIED	02:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:47
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:48
BLANK TEST	0.00	02:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:50
BLANK TEST	0.00	02:51

