

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/01/2024

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:14
INTERNAL STANDARD	VERIFIED	05:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:15
BLANK TEST	0.00	05:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:16
BLANK TEST	0.00	05:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:17
BLANK TEST	0.00	05:18

