

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/08/2024

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:06
BLANK TEST	0.00	02:06

